



180015144061



中国认可
国际互认
检测
TESTING
CNAS L2954

Final Report

Report Number: SDWH-M202101574-2(E)

Acute Systemic Toxicity of Disposable Nitrile Examination Gloves

According to ISO 10993-11:2017
Intraperitoneal
Sesame Oil Extract

Sponsor: Shandong Intco Medical Products Co.,Ltd

Address: No.9888,Qiwang Road,Naoshan Industry
Park,Qingzhou,Shandong,China



Sanitation & Environment Technology Institute, Soochow University

Address: 199 Ren-Ai Road, Suzhou Industrial Park, Suzhou, Jiangsu 215123, P. R. China

Website: www.sudatest.com

E-mail: med@sudatest.com

Direct: +86 512 65880038

Free: 400 107 8828

Content

| | |
|---|-----------|
| Supplementary Explanation | 3 |
| Verification Dates | 4 |
| Summary..... | 5 |
| Test Report..... | 6 |
| 1 Purpose..... | 6 |
| 2 Reference | 6 |
| 3 Compliance | 6 |
| 4 Identification of Test and Control Articles | 6 |
| 4.1 Test Article | 6 |
| 4.2 Control Article..... | 7 |
| 4.2.1 Negative Control | 7 |
| 5 Equipment and Reagents | 7 |
| 5.1 Equipment | 7 |
| 5.2 Reagents..... | 7 |
| 6 Identification of Test System | 7 |
| 7 Animal Care and Maintenance..... | 7 |
| 8 Justification of Test System and Route of Administration | 8 |
| 9 Experimental Design..... | 8 |
| 9.1 Preparation of Extracts..... | 8 |
| 9.1.1 Pretreatment | 8 |
| 9.1.2 Extraction | 8 |
| 9.2 Experimental Procedure..... | 8 |
| 9.2.1 Animal Preparation and Grouping..... | 8 |
| 9.3 Evaluation of Results | 9 |
| 10 Results | 9 |
| 11 Conclusion | 9 |
| 12 Record Storage | 9 |
| 13 Confidentiality Agreement | 9 |
| 14 Deviation Statement..... | 9 |
| Annex 1 Test Data | 10 |
| Annex 2 Photograph of Test Article | 11 |
| Annex 3 Information Provided by Sponsor..... | 12 |

Supplementary Explanation

- (1) Please apply for rechecking within 15 days of receiving the report if there are any objections.
- (2) Any erasure or without special inspection and testing seal renders the report null and void.
- (3) The report is only valid when signed by the persons who edited, checked and approved it.
- (4) The results relate only to the articles tested.
- (5) The report shall not be reproduced except in full without the written approval of the institute.

Verification Dates

| | |
|------------------------------|------------|
| Test Article Receipt | 2021-04-01 |
| Protocol Effective Date | 2021-04-09 |
| Technical Initiation Date | 2021-04-09 |
| Technical Completion Date | 2021-04-16 |
| Final Report Completion Date | 2021-04-28 |

Edited by: Wang Deheng 2021-04-25
Date

Reviewed by: Lu Chenxi 2021-04-28
Study Director Date

Approved by: Fang Jingyi 2021-04-28
Authorized Signatory Date

Sanitation & Environment Technology Institute, Soochow University



Summary

1 Test Article

| | |
|--------------------------|--|
| Test Article Name | Disposable Nitrile Examination Gloves |
| Manufacturer | Shandong Intco Medical Products Co.,Ltd |
| Address | No.9888,Qiwan Road,Naoshan Industry Park,Qingzhou,Shandong,China |
| Model | N/A |
| Lot/Batch | N/A |

2 Main Reference

ISO 10993-11:2017 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity

3 Test Method

The test article was extracted and the extract was evaluated to determine whether leachables extracted from the test article would cause acute systemic toxicity following injection into mice test in accordance with ISO 10993-11:2017 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity.

Study protocol number: SDWH-PROTOCOL- M202101574-2.

4 Conclusion

Under the conditions of this study, there was no evidence of systemic toxicity from the extract. The test article extract met the requirements of this study.

Test Report

1 Purpose

The test article was extracted and the extract was evaluated to determine whether leachables extracted from the test article would cause acute systemic toxicity following injection into mice.

2 Reference

ISO 10993-11:2017 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity

ISO 10993-12:2021 Biological evaluation of medical devices — Part 12: Sample preparation and reference materials

ISO 10993-2:2006 Biological evaluation of medical devices — Part 2: Animal welfare requirements

3 Compliance

ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories (CNAS—CL01 Accreditation criteria for the competence of testing and calibration laboratories) China National Accreditation Service for Conformity Assessment LABORATORY ACCREDITATION CERTIFICATE Registration No. CNAS L2954.

RB/T 214—2017 Competence assessment for inspection body and laboratory mandatory approval—General requirements for inspection body and laboratory Certification and Accreditation Administration of the People's Republic of China INSPECTION BODY AND LABORATORY MANDATORY APPROVAL Certificate No. CMA 180015144061.

4 Identification of Test and Control Articles

4.1 Test Article

| | |
|----------------------------|---|
| Test Article Name | Disposable Nitrile Examination Gloves |
| Manufacturer | Shandong Intco Medical Products Co.,Ltd |
| Address | No.9888,Qiwang Road,Naoshan Industry Park, Qingzhou,Shandong,China |
| Test Article Initial State | Non-sterile |
| CAS Number | Not supplied by sponsor (N/S) |
| Model | N/A |
| Size | N/A |
| Lot/Batch | N/A |
| Raw Material | Nitrile Latex (NBR) |
| Packaging Material | N/A |
| Physical State | Solid |
| Color | Blue |
| Density | N/A |
| Stability | N/A |
| Solubility | N/A |
| Storage Condition | Room temperature |
| Intended Use | It is mainly used for Patient Examination, to prevent possible cross infection between users and patients and prevent cross infection between patients. |
| Additional Information | N/A |

The information about the test article was supplied by the sponsor wherever applicable.

4.2 Control Article

4.2.1 Negative Control

Article Name: Sesame Oil (SO)

Manufacturer: Ji'an Qingyuan District luyuanxiangliao. Co. Ltd

Size: 5kg

Lot/ Batch#: 20201218

Physical State: Liquid

Color: Light yellow

Storage Condition: Room Temperature

5 Equipment and Reagents

5.1 Equipment

| Equipment Name | Equipment Number | Calibration Expire |
|---|------------------|--------------------|
| Horizontal Large Capacity Constant Temperature Vibrator | SDWH2718 | 2021-08-10 |
| Vertical pressure steam sterilizer | SDWH2097 | 2022-03-09 |
| Steel straight scale | SDWH463 | 2021-07-06 |
| Electronic scale | SDWH131 | 2021-12-08 |

5.2 Reagents

| Reagent Name | Manufacturer | LOT |
|-----------------|---|----------|
| Sesame oil (SO) | Ji'an Qingyuan District luyuanxiangliao Co., Ltd. | 20201218 |

6 Identification of Test System

Species: ICR mouse

Number: 10

Sex: Males

Initial body weight: Initial weight 18-22 g. The weight variation of animals used within a sex did not exceed $\pm 20\%$ of the mean weight.

Health status: Healthy, Young

Housing: Five animals per cage identified by a card indicating the lab number, test code, and first treatment date, etc.

Animal identification: Stain with picric acid

Cages: Plastic cages

Acclimation period: 7 days under the same conditions as for the actual test

7 Animal Care and Maintenance

Animal source: Provided by Zhejiang Vital River Laboratory Animal Technology Co. Ltd.; Permit Code SCXK (Zhe) 2019-0001

Bedding: Corncob (Suzhou shuangshi laboratory animal feed science Co., Ltd)

Feed: Irradiation sterilization feed (Suzhou shuangshi laboratory animal feed science Co., Ltd)

Water: Sterile water

Animal room temperature: 20-26°C

Animal room relative humidity: 30%-70%

Lights: 12 hours light/dark cycle, full spectrum lights

Personnel: Associates involved were appropriately qualified and trained.

Selection: Only healthy, previously unused animals were selected.

Device sterilization: All devices which will contact with the testing sample are subjected to steam sterilization in an autoclave at 121°C for 30min.

There were no known contaminants present in the feed, water, or bedding expected to interfere with the test data.

8 Justification of Test System and Route of Administration

Mice have historically been used in biological evaluation studies. The species and number of animals as well as the route of administration used are specified in the current standard for evaluation of medical devices.

The test article was exposed to the test system through a solvent compatible with the test system following intraperitoneal administration (IP). This was the optimal route of administration available in this test system.

9 Experimental Design

9.1 Preparation of Extracts

9.1.1 Pretreatment

No pretreatment required.

9.1.2 Extraction

Under aseptic conditions, samples were taken according to the sampling method (Random sampling, calculating the combined area of all tissue contacting surfaces of each sample (one side) as the standard surface area of each sample). The extraction was performed with agitation in closed inert containers according to the extraction ratio listed in the following table(sample: extraction vehicle). The extraction vehicle was SO.

| Samples | Actual Sampling | Extract Procedure | | | Final Extract |
|------------------|--------------------------------------|--------------------------|---------|------------|---------------|
| | | Extract Ratio | SO | Condition | |
| Test | Surface area 60.0 cm ² | 6 cm ² : 1 mL | 10.0 mL | 50°C, 72 h | Clear |
| Negative Control | / | / | 20.0 mL | 50°C, 72 h | Clear |

The state of the extract did not change after extraction. The extract was stored at room temperature, and tested within 24 h, without the process of adjusting its pH value, filtering, centrifuging, diluting, etc.

The vehicle (without the test article) was similarly prepared to serve as the control.

9.2 Experimental Procedure

9.2.1 Animal Preparation and Grouping

Prior to dosing, the mice were individually identified, weighed, and randomly assigned to a treatment group as shown below:

| Group Name | Group Size | Sex | Dose Level | Route |
|------------------|------------|------|------------|-----------------|
| Test | 5 animals | Male | 50 mL/kg | Intraperitoneal |
| Negative Control | 5 animals | Male | 50 mL/kg | Intraperitoneal |

A single dose of test article extract was injected into the designated group of mice intraperitoneally at the dose level of 50 mL/kg bw. The negative control liquid was injected similarly into the separate group of designated control mice.

Mice were observed for any adverse clinical reactions immediately after injection, and then the animals were returned to their cages. The animals were observed for signs of systemic reactions at 4, 24, 48 and 72 hours after injection and weighed daily for three days after dosing. Any animal found dead or showed abnormal signs were subjected to gross necropsy.

9.3 Evaluation of Results

(1) If during the observation period of an acute systemic toxicity test none of the mice treated with the test article extract exhibited a significantly greater biological reactivity than control mice, the test article met the requirements. If two or more animals died, or if abnormal behavior such as convulsions or prostration occurs in two or more animals, or if body weight loss greater than 10 % occurs in three or more animals, the test article did not meet the requirements.

(2) If any animals treated with the sample exhibited only slight signs of biological reactivity, and no more than one animal showed gross symptoms of biological reactivity or died, repeat the testing using groups of ten animals. On the repeat test, if all ten animals treated with the test article extract exhibited no scientifically meaningful biological reactivity above the vehicle control animals during the observation period, the test article met the requirements.

10 Results

Clinical Observations

All animals appeared clinical normal throughout the study. The clinical observations are presented in Table 1 in the appendix.

Body Weight

Body weight data were acceptable and equivalent between the corresponding test and control treatment groups. Body weight data are presented in Table 2 in the appendix.

11 Conclusion

Under the conditions of this study, there was no evidence of systemic toxicity from the extract. The test article extract met the requirements of the study.

12 Record Storage

All the raw data and records pertaining to this study were retained in designated SDWH archive.

13 Confidentiality Agreement

Statements of confidentiality were as agreed upon prior to study initiation.

14 Deviation Statement

There were no deviations from the approved study protocol which were judged to have any impact on the validity of the data.

Annex 1 Test Data

Table 1 Clinical Observations

| Treatment Group | Animal Number | Observation (after injection) | | | | |
|-----------------|---------------|-------------------------------|--------------------|--------------------|--------------------|--------------------|
| | | Immediate | 4 Hours | 24 Hours | 48 Hours | 72 Hours |
| Test | 1 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 2 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 3 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 4 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 5 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| Control | 6 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 7 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 8 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 9 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |
| | 10 | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal | Appeared Normal |

Table 2 Body Weight Data

| Treatment Group | Animal Number | Initial Weight (g) | Weight after Injection (g) | | | |
|-----------------|---------------|--------------------|----------------------------|----------|----------|----------|
| | | | 4 Hours | 24 Hours | 48 Hours | 72 Hours |
| Test | 1 | 18.4 | 18.5 | 18.7 | 19.2 | 20.1 |
| | 2 | 20.4 | 21.0 | 20.6 | 21.1 | 22.0 |
| | 3 | 19.6 | 19.9 | 20.6 | 21.7 | 22.8 |
| | 4 | 19.5 | 20.0 | 20.0 | 20.5 | 21.7 |
| | 5 | 19.2 | 19.8 | 20.2 | 21.3 | 22.2 |
| Control | 6 | 19.2 | 19.5 | 19.9 | 20.5 | 21.6 |
| | 7 | 20.5 | 20.9 | 21.5 | 22.0 | 22.4 |
| | 8 | 20.7 | 21.4 | 21.4 | 22.4 | 23.2 |
| | 9 | 19.5 | 19.4 | 20.6 | 21.0 | 21.8 |
| | 10 | 20.4 | 20.9 | 21.0 | 22.0 | 22.9 |

Annex 2 Photograph of Test Article



Annex 3 Information Provided by Sponsor

1 Production Process

Not supplied by sponsor.

2 Other Information

Not supplied by sponsor.

End of Report